

ACE - CONTROL SET

Animal Control sera of Angiotensin converting enzyme for the control of dosage method in serum

INTENDED USE

Concentrations and activities of components are studied for an excellent use even in automatic system of clinical chemistry, even in manual procedures.

PRESERVATION AND STABILITY

Preserve the product at 2-8°C, stable until expiration date on the package.

Once opened controls are stable for 8 hours at 2-8°C or 15 days at -20°C.

MATERIALS REQUIRED BUT NOT SUPPLIED

Automatic Micropipette, normal laboratory equipment, distilled or deionized water.

REAGENTS PREPARATIONS

Open carefully one vial of liophile, trying to avoid loss of material, and with the enclosed pipette add exactly 1.0 ml of deionized water, for this is better employing an automatic pipette.

Close with attention and let the liophile recovers for 30 minutes, shake carefully trying to avoid foam formation.

Bring controls at room temperature before use.

PRECAUTIONS

- Product contains Sodium Azide 0.095% as preservative.
- Avoid contact with skin and mucous membrane.
- Use the normal precautions for lab behaviour.
- For in vitro use only.

WASTE DISPOSAL

Product is intended for professional laboratories. Waste products must be handled as per relevant security cards and local regulations.

PACKAGING

CODE CC03930

Normal Control 1 x 1 ml (liophile)
Pathological Control 1 x 1 ml (liophile)

REFERENCES



Archives data.

CALIBRATORS VALUES

LOT G044.26*

 2028/02

This Table of valueset for lot is applicable to sublots identified by the sequential alphabetical letter (for example A, B, C, etc.) following the lot number.

COMPONENT / METHOD		VALUE	RANGE	
NORMAL CONTROL	Ronca-Testoni (LTA Reagents)	98.7	80.8 – 116.3	U/I
	LOT R033.26A  2028/02 Lieberman's method**	60.2	49.4 – 71.03	U/I
PATHOLOGICAL CONTROL	Ronca-Testoni (LTA Reagents)	194.4	159.4 – 229.4	U/I
	LOT S033.26A  2028/02 Lieberman's method**	118.6	97.2 – 139.9	U/I






****Other commercially kits for the determination of ACE, using the same method, using a different factor calculation.**

In order to compare the results obtained from the LTA kit with those obtained from other kits, you must multiply by 0.61.

MANUFACTURER

LTA s.r.l.
Via Milano 15/F
20060 Bussero (Milan) ITALY
Tel: +39 02 95409034
Fax: +39 02 95334185
e-mail: info@LTAonline.it
Website: http://www.LTAonline.it

SYMBOLS

- IVD** Only for IVD use
- LOT** Lot of manufacturing
- REF** Code number
-  Storage temperature interval
-  Expiration date (year – month)
-  Warning, read enclosed documents
-  Read the directions
-  Biological risk

Mod. 01.06 (ver. 1.5 - 09/05/2018)

