

# ACE CALIBRATOR

**Animal Calibration sera of Angiotensin converting enzyme for the control of dosage method in serum**

## INTENDED USE

Concentrations and activities of components are studied for an excellent use even in automatic system of clinical chemistry, even in manual procedures.

## PRESERVATION AND STABILITY

Preserve the product not reconstituted at 2-8°C, stable until expiration date on the package.  
Calibrator reconstituted is stable 8 hour at 2-8°C or 15 days at -20°C.

## MATERIALS REQUIRED BUT NOT SUPPLIED

Automatic Micropipette, normal laboratory equipment, distilled or deionised water.

## REAGENTS PREPARATION

Open carefully one vial of liophile, trying to avoid loss of material, and with the enclosed pipette add exactly 1.0 ml of distilled water, for this is better employing an automatic pipette. Close with attention and let the liophile recovers for 30 minutes, shake carefully trying to avoid foam formation.

## PRECAUTIONS

- Product contains Sodium Azide 0.095% as preservative.
- Avoid contact with skin and mucous membrane.
- Use the normal precautions for lab behaviour.
- For in vitro use only.

## WASTE DISPOSAL

Product is intended for professional laboratories. Waste products must be handled as per relevant security cards and local regulations.

## PACKAGING

### CODE CC03940


Calibrator 2 x 1 ml (liophile)

## REFERENCES

Archives data.

## CALIBRATORS VALUES

**LOT** F300.25A\*

 2027/10

*\*This Table of value set for lot is applicable to sublots identified by the sequential alphabetical letter (for example A, B, C, etc.) following the lot number.*









METHOD	CALIBRATION VALUE	UNIT
Ronca-Testoni (LTA Reagents)	143.4	U/l
Lieberman's method**	86.8	U/l

**\*\*Other commercially kits for the determination of ACE, using the same method, using a different factor calculation.  
In order to compare the results obtained from the LTA kit with those obtained from other kits, you must multiply by 0.61.**

## MANUFACTURER

LTA s.r.l.  
Via Milano 15/F  
20060 Bussero (Milan) ITALY  
Tel: +39 02 95409034  
Fax: +39 02 95334185  
e-mail: info@LTAonline.it  
Website: <http://www.LTAonline.it>

## SYMBOLS

-  Only for IVD use
-  Lot of manufacturing
-  Code number
-  Storage temperature interval
-  Expiration date
-  Warning, read enclosed documents
-  Read the directions
-  Biological risk

Mod. 01.06 (ver. 1.2 – 09/05/2018)

